

4^{7/8}" (12.4 cm)



3" (7.6 cm)

**Recommended Clearance
for Imager**

4^{7/8}" (12.4 cm)

4^{1/4}" (10.8 cm)

**Minimum Required
Clearance for Imager
Display**

4^{3/8}" (11 cm)



7^{7/8}" (19.8 cm)



4^{3/8}" (11 cm)

**8" (20.3 cm) Clearance Recommended
for Optional Accessory Storage**

60510011279B (1116)